

Manual Supplement

Manual Title:	DS701, DS703 FC Users	Supplement Issue:	3
Print Date:	March 2018	Date:	8/18
Revision/Date:		Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 575

For the DS703 FC only:

Fluke Connect functionality is coming soon. Go to <http://a.fluke.com/registration> and register your product to receive notification when Fluke Connect is available in your region.

Change #2, 632, 635

On page 19, under **Specifications**, replace **Temperature** section with:

Operating	0 °C to +45 °C (+32 °F to +113 °F)
Storage	-10 °C to +60 °C (-14 °F to +140 °F)
Battery Charging	0 °C to 35 °C (+32 °F to +95 °F)
Rating	5 V DC, 1.3 A

